

Sec. 21a-63-3. Factory records

Each permittee shall keep on file for at least two years complete records of each clinical thermometer which has been sold in the state of Connecticut, the record to include either a serial number or code which indicates the specific period, not to exceed 90 days, in which the thermometers were calibrated, name and address of the purchaser, and the date of sale of each lot of thermometers sold. These records shall be available to a representative of the State Department of Consumer Protection at any time upon request.

(Effective February 29, 1988)