

**Sec. 21a-63-12. Test for entrapped gas**

Gas in bulb. Thermometers in which inspection shows the presence of gas in the bulb shall be rejected.

(Effective July 27, 1984)